

Integrated Biomedical Technology, Inc.**Document No.**

Elkhart, Indiana 46514

QA32-221

Phone: 574-264-0025 Toll Free: 1-800-490-5500 Fax: 574-264-2787

Certificate of Analysis**Product: WaterCheck RC (RC22101)**Lot No. 11229021Exp. Date: 03/12

Chlorine Control Solutions	Color Blocks (Pantone #)	Strip Color	Tested
500 ppm Chlorine	Black Brown (532C)	<u>500 ppm color</u>	<u>Pass</u>
1.0 ppm Chlorine	Ocean Blue (629C)	<u>1.0 ppm color</u>	<u>Pass</u>
0.5 ppm Chlorine	Positive (628C)	<u>0.5 ppm color</u>	<u>Pass</u>
H ₂ O DI water	Negative (White)	<u>Off white</u>	<u>Pass</u>

Approved by



Lia Harrison

Title

QA Chemist

Date

9/14/09

This analysis was done at the manufacturing as part of the Quality Assurance Program in accordance with the FDA Quality System Regulation. The State and the local JCAHO inspectors may also require the end user of the product to perform a quality control check before using the strips. The end user is advised to follow the quality control procedure as described in the product insert.